



EPA METHOD 8270

For the analysis of semi-volatiles
in solid waste and wastewater

lgcstandards.com/foodandenvironment

LGC Quality | ISO 9001 | ISO/IEC 17025 | ISO Guide 34 |

WE'RE MEETING EPA METHOD 8270 STANDARDS

For semi-volatiles in solid
waste and wastewater

Dr. Ehrenstorfer offers the full range of easy-to-use mixes to test semi-volatiles in solid waste and wastewater in accordance with EPA Method 8270.

Designed for convenience, precision and accuracy, the product group features Dr. Ehrenstorfer's new 76 component all-in-one 8270 UltiMix for instrument calibration. Reactive analytes have been removed from the mix and are offered separately to make calibration and analysis easier. The mixtures in this group are manufactured in accordance with ISO Guide 34 requirements and are part of our Dr. Ehrenstorfer product line which offers thousands of reference standards for residue analysis.


If none of our stock standards suit your needs, we will manufacture a custom standard to your specifications.

Please let us know if you have questions or would like to discuss your specific requirements.

We are here to help.



Email: lgcusa@lgcstandards.com
Order online at: lgcstandards.com
Prefer to speak to someone
in the US?: +1 603 622 7660
Additional sales office contact
details are listed at the back.



CONTENTS

Table of contents

Introduction

Calibration mixes

UltiMix 8270 Calibration Mix (76 components)	1
Benzidines Mix	2
Organochlorine Mix (21 components)	2

Calibration mixes by class


Calibration Mix 1 (22 components)	2
Calibration Mix 2 (11 components)	3
Calibration Mix 3 (23 components)	3
Calibration Mix 4 (22 components)	3
Calibration Mix 5 (19 components)	4
Calibration Mix 6 (10 components)	4
EPA Method 8270 Appendix IX Mix 1 with Phentermine (17 components)	4
EPA Method 8270 Appendix IX Mix 1 minus Phentermine (16 components)	5
EPA Method 8270 Appendix IX Mix 2 (31 components)	5
EPA Method 8270 Appendix IX Organophosphate Mix (9 components)	5
Aramite Solution	6
Methapyrilene Solution	6

Quality control products

EPA Method 8270 UltiMix Matrix Spike Mix (76 components)	6
Semi-volatile (BNA) Matrix Spike Mix (11 components)	7
EPA Method 8270 B/N Calibration Check Mix (7 components)	7
EPA Method 8270 Acid Calibration Check Mix (6 components)	7
EPA Method 8270 Internal Standard Mix (7 components)	7
Color Changing BNA Surrogates (7 components 25 mL)	8
BNA Surrogates (6 components)	8
BNA Surrogates (7 components)	8
EPA Method 8270 Semi-volatile System Performance Check Mix (4 components)	9

Tuning

PFTBA MS Tuning Compound	9
GC Tuning Mix (4 components)	9
DFTPP SV Tuning Compound	9



EPA METHOD 8270

Calibration mixes

UltiMix 8270 Calibration Mix (76 components)		Product Number	Size
Reactive analytes offered separately 500/1000 µg/mL in Methylene Chloride/Benzene 1:1 3-methylphenol and 4-methylphenol are 500 µg/mL each		DRE-GA09000419BD	1 mL
acenaphthene	83-32-9	2,6-dinitrotoluene	606-20-2
acenaphthylene	208-96-8	di-n-octyl phthalate	117-84-0
aniline	62-53-3	diphenylamine	122-39-4
anthracene	120-12-7	fluoranthene	206-44-0
azobenzene	103-33-3	fluorene	86-73-7
benzo[a]anthracene	56-55-3	hexachlorobenzene	118-74-1
benzo[b]fluoranthene	205-99-2	hexachlorobutadiene	87-68-3
benzo[k]fluoranthene	207-08-9	hexachlorocyclopentadiene	77-47-4
benzo[ghi]perylene	191-24-2	hexachloroethane	67-72-1
benzo[a]pyrene	50-32-8	indeno[1,2,3-cd]pyrene	193-39-5
benzyl alcohol	100-51-6	isophorone	78-59-1
bis(2-chloroethoxy)methane	111-91-1	2-methyl-4,6-dinitrophenol (wet)	534-52-1
bis(2-chloroethyl)ether	111-44-4	1-methylnaphthalene	90-12-0
bis(2-chloro-1-methylethyl) ether	108-60-1	2-methylnaphthalene	91-57-6
4-bromophenyl phenyl ether	101-55-3	2-methylphenol	95-48-7
butyl benzyl phthalate	85-68-7	3-methylphenol	108-39-4
carbazole	86-74-8	4-methylphenol	106-44-5
4-chloroaniline	106-47-8	naphthalene	91-20-3
4-chlorodiphenyl ether	7005-72-3	2-nitroaniline	88-74-4
4-chloro-3-methylphenol	59-50-7	3-nitroaniline	99-09-2
2-chloronaphthalene	91-58-7	4-nitroaniline	100-01-6
2-chlorophenol	95-57-8	nitrobenzene	98-95-3
chrysene	218-01-9	2-nitrophenol	88-75-5
dibenz[a,h]anthracene	53-70-3	4-nitrophenol	100-02-7
dibenzofuran	132-64-9	N-nitrosodimethylamine	62-75-9
di-n-butyl phthalate	84-74-2	N-nitrosodi-n-propylamine	621-64-7
1,2-dichlorobenzene	95-50-1	pentachlorophenol	87-86-5
1,3-dichlorobenzene	541-73-1	phenanthrene	85-01-8
1,4-dichlorobenzene	106-46-7	phenol	108-95-2
2,4-dichlorophenol	120-83-2	pyrene	129-00-0
diethyl phthalate	84-66-2	pyridine	110-86-1
2,4-dimethylphenol	105-67-9	2,3,4,6-tetrachlorophenol	58-90-2
dimethyl phthalate	131-11-3	2,3,5,6-tetrachlorophenol	935-95-5
1,2-dinitrobenzene	528-29-0	1,2,4-trichlorobenzene	120-82-1
1,3-dinitrobenzene	99-65-0	2,4,5-trichlorophenol	95-95-4
1,4-dinitrobenzene	100-25-4	2,4,6-trichlorophenol	88-06-2
2,4-dinitrophenol	51-28-5	bis(2-ethylhexyl) adipate	103-23-1
2,4-dinitrotoluene	121-14-2	bis(2-ethylhexyl) phthalate	117-81-7

*1,2-diphenylhydrazine decomposes to azobenzene in the injector and is not included.

*N-nitrosodiphenylamine decomposes to diphenylamine in the injector and is not included.

EPA METHOD 8270

Calibration mixes

Benzidines Mix		Product Number	Size
2000 µg/mL in Methanol		DRE-GA09000426ME	1 mL
		DRE-GS09000426ME	5 x 1 mL
benzidine	92-87-5		
3,3'-dichlorobenzidine	91-94-1		
3,3'-dimethylbenzidine	119-93-7		

Organochlorine Mix (21 components)		Product Number	Size
1000 µg/mL in MtBE		DRE-GA09000427MB	1 mL
		DRE-GS09000427MB	5 x 1 mL
aldrin	309-00-2	endosulfan II	33213-65-9
a-BHC	319-84-6	endosulfan sulfate	1031-07-8
b-BHC	319-85-7	endrin	72-20-8
d-BHC	319-86-8	endrin aldehyde	7421-93-4
g-BHC	58-89-9	endrin ketone	53494-70-5
cis-chlordane	5103-71-9	heptachlor	76-44-8
p,p'-DDD	72-54-8	heptachlor epoxide isomer B	1024-57-3
p,p'-DDE	72-55-9	hydroquinone	123-31-9
p,p'-DDT	50-29-3	methoxychlor	72-43-5
dieldrin	60-57-1	trans-chlordane	5103-74-2
endosulfan I	959-98-8		

Calibration mixes by class

Calibration Mix 1 (22 components)		Product Number	Size
2000 µg/mL in Methylene Chloride		DRE-GA09000437DI	1 mL
benzoic acid	65-85-0	2-methylphenol	95-48-7
4-chloro-3-methylphenol	59-50-7	3-methylphenol	108-39-4
2-chlorophenol	95-57-8	4-methylphenol	106-44-5
2,4-dichlorophenol	120-83-2	2-nitrophenol	88-75-5
2,6-dichlorophenol	87-65-0	4-nitrophenol	100-02-7
2,4-dimethylphenol	105-67-9	pentachlorophenol	87-86-5
2,4-dinitrophenol	51-28-5	phenol	108-95-2
dinoseb	88-85-7	2,3,4,6-tetrachlorophenol	58-90-2
2-fluorophenol	367-12-4	2,4,6-tribromophenol	118-79-6
hexachlorophene	70-30-4	2,4,5-trichlorophenol	95-95-4
2-methyl-4,6-dinitrophenol	534-52-1	2,4,6-trichlorophenol	88-06-2

EPA METHOD 8270

Calibration mixes by class

Calibration Mix 2 (11 components)		Product Number	Size
Manufactured upon request* 2000 µg/mL in Methylene Chloride		DRE-GA09000438DI	1 mL
aniline	62-53-3	3-nitroaniline	99-09-2
benzidine	92-87-5	4-nitroaniline	100-01-6
4-chloroaniline	106-47-8	N-nitrosodimethylamine	62-75-9
3,3'-dichlorobenzidine	91-94-1	N-nitrosodi-n-propylamine	621-64-7
diphenylamine	122-39-4	pyridine	110-86-1
2-nitroaniline	88-74-4		

*Even with our state of the art analyte stabilization, some products have shorter shelf lives and are manufactured upon request.

Calibration Mix 3 (23 components)		Product Number	Size
2000 µg/mL in Methylene Chloride		DRE-GA09000439DI	1 mL
aramite	140-57-8	hexachlorobutadiene	87-68-3
bis(2-chloroethoxy) methane	111-91-1	hexachlorocyclopentadiene	77-47-4
bis(2-chloroethyl) ether	111-44-4	hexachloroethane	67-72-1
bis(2-chloro-1-methylethyl) ether	108-60-1	hexachloropropene	1888-71-7
chlorobenzilate	510-15-6	isodrin	465-73-6
4-chlorodiphenyl ether	7005-72-3	kepone	143-50-0
2-chloronaphthalene	91-58-7	pentachlorobenzene	608-93-5
1,2-dichlorobenzene	95-50-1	pentachloronitrobenzene	82-68-8
1,3-dichlorobenzene	541-73-1	1,2,4,5-tetrachlorobenzene	95-94-3
1,4-dichlorobenzene	106-46-7	1,2,4-trichlorobenzene	120-82-1
1,3-dinitrobenzene	99-65-0	4-bromophenylpheny ether	101-55-3
hexachlorobenzene	118-74-1		

Calibration Mix 4 (22 components)		Product Number	Size
2000 µg/mL in Methylene Chloride		DRE-GA09000440DI	1 mL
acetophenone	98-86-2	isophorone	78-59-1
azobenzene	103-33-3	isosafrole	120-58-1
benzyl alcohol	100-51-6	methyl methanesulfonate	66-27-3
butyl benzyl phthalate	85-68-7	1,4-naphthoquinone	130-15-4
dibenzofuran	132-64-9	nitrobenzene	98-95-3
di-n-butyl phthalate	84-74-2	4-nitroquinoline-1-oxide	56-57-5
diethyl phthalate	84-66-2	phenacetin	62-44-2
dimethyl phthalate	131-11-3	safrole	94-59-7
2,4-dinitrotoluene	121-14-2	1,3,5-trinitrobenzene	99-35-4
2,6-dinitrotoluene	606-20-2	dioctyl phthalate	117-81-7
di-n-octyl phthalate	117-84-0	ethyl methanesulfonate	62-50-0

EPA METHOD 8270

Calibration mixes by class

Calibration Mix 5 (19 components)		Product Number	Size
2000 µg/mL in Methylene Chloride		DRE-GA09000441DI	1 mL
acenaphthene	83-32-9	fluoranthene	206-44-0
acenaphthylene	208-96-8	fluorene	86-73-7
anthracene	120-12-7	indeno[1,2,3-cd]pyrene	193-39-5
benzo[a]anthracene	56-55-3	3-methylcholanthrene	56-49-5
benzo[b]fluoranthene	205-99-2	1-methylnaphthalene	90-12-0
benzo[k]fluoranthene	207-08-9	2-methylnaphthalene	91-57-6
benzo[ghi]perylene	191-24-2	naphthalene	91-20-3
benzo[a]pyrene	50-32-8	phenanthrene	85-01-8
chrysene	218-01-9	pyrene	129-00-0
dibenz[a,h]anthracene	53-70-3		

Calibration Mix 6 (10 components)		Product Number	Size
2000 µg/mL in Methylene Chloride		DRE-GA09000442DI	1 mL
di-allate (mixture of isomers)	2303-16-4	parathion	56-38-2
dimethoate	60-51-5	phorate	298-02-2
disulfoton	298-04-4	propyzamide (pronamide)	23950-58-5
famphur	52-85-7	thionazine (zinophos)	297-97-2
methyl parathion	298-00-0	O,O,O-triethylphosphorothioate	126-68-1

Appendix IX Mixes

EPA Method 8270 Appendix IX Mix 1 with Phentermine (17 components)		Product Number	Size
Benzidine and other reactive compounds offered separately			
2000 µg/mL in Methylene Chloride		DRE-GA09000443DI	1 mL
2-acetylaminofluorene	53-96-3	N-nitrosodi-n-butylamine	924-16-3
4-aminobiphenyl	92-67-1	N-nitrosomethylethylamine	10595-95-6
p-(dimethylamino) azobenzene	60-11-7	N-nitrosomorpholine	59-89-2
phentermine	122-09-8	N-nitrosopiperidine	100-75-4
isosafrole	120-58-1	N-nitrosopyrrolidine	930-55-2
1-naphthylamine	134-32-7	p-phenylenediamine	106-50-3
2-naphthylamine	91-59-8	2-picoline	109-06-8
5-nitro-o-toluidine	99-55-8	o-toluidine	95-53-4
n-nitrosodiethylamine	55-18-5		

EPA METHOD 8270

Appendix IX Mixes

EPA Method 8270 Appendix IX Mix 1 minus Phentermine (16 components)		Product Number	Size
Benzidine and other reactive compounds offered separately 2000 µg/mL in Methylene Chloride		DRE-GA09000444DI	1 mL
		DRE-GS09000444DI	5 x 1 mL
2-acetylaminofluorene	53-96-3	n-nitrosodi-n-butylamine	924-16-3
4-aminobiphenyl	92-67-1	N-nitrosomethylethylamine	10595-95-6
p-(dimethylamino) azobenzene	60-11-7	N-nitrosomorpholine	59-89-2
isosafrole	120-58-1	N-nitrosopiperidine	100-75-4
1-naphthylamine	134-32-7	N-nitrosopyrrolidine	930-55-2
2-naphthylamine	91-59-8	p-phenylenediamine	106-50-3
5-nitro-o-toluidine	99-55-8	2-picoline	109-06-8
n-nitrosodiethylamine	55-18-5	o-toluidine	95-53-4

EPA Method 8270 Appendix IX Mix 2 (31 components)		Product Number	Size
Manufactured upon request 1000 µg/mL in Methylene Chloride		DRE-GA09000445DI	1 mL
		DRE-GS09000445DI	5 x 1 mL
acetophenone	98-86-2	isodrin	465-73-6
aramite	140-57-8	isosafrole	120-58-1
benzaldehyde	100-52-7	kepone	143-50-0
biphenyl	92-52-4	3-methylcholanthrene	56-49-5
caprolactam	105-60-2	methyl methanesulfonate	66-27-3
chlorobenzilate	510-15-6	1,4-naphthoquinone	130-15-4
1-chloronaphthalene	90-13-1	4-nitroquinoline-1-oxide	56-57-5
di-allate (mixture of isomers)	2303-16-4	pentachlorobenzene	608-93-5
dibenz(a,j) acridine	224-42-0	pentachloroethane	76-01-7
2,6-dichlorophenol	87-65-0	pentachloronitrobenzene	82-68-8
7,12-dimethylbenz[a]anthracene	57-97-6	phenacetin	62-44-2
1,4-dioxane	123-91-1	propyzamide (pronamide)	23950-58-5
diphenyl ether	101-84-8	safrole	94-59-7
ethyl methacrylate	97-63-2	1,2,4,5-tetrachlorobenzene	95-94-3
ethyl methanesulfonate	62-50-0	1,3,5-trinitrobenzene	99-35-4
hexachloropropene	1888-71-7		

EPA Method 8270 Appendix IX Organophosphate Mix (9 components)		Product Number	Size
2000 µg/mL in Methylene Chloride		DRE-GA09000364DI	1 mL
dimethoate	60-51-5	phorate	298-02-2
disulfoton	298-04-4	thionazine (zinophos)	297-97-2
famphur	52-85-7	O,O,O-triethylphosphorothioate	126-68-1
methyl parathion	298-00-0	tetraethyldithio pyrophosphate	3689-24-5
parathion	56-38-2		

EPA METHOD 8270

Aramite Solution		Product Number	Size
2000 µg/mL in Hexane	140-57-8	DRE-GA09010391HE	1 mL

Methapyrilene Solution		Product Number	Size
2000 µg/mL in Methylene Chloride	91-80-5	DRE-GA09010392DI	1 mL
		DRE-GS09010392DI	5 x 1 mL

Quality control products

EPA Method 8270 UltiMix Matrix Spike Mix (76 components)		Product Number	Size
3-methylphenol and 4-methylphenol are 100 µg/mL each 100/200 µg/mL in Methanol/Methylene Chloride 80:20		DRE-GA09000420MD	5 mL
		DRE-GA09000421MD	2 x 5 mL
acenaphthene	83-32-9	2,6-dinitrotoluene	606-20-2
acenaphthylene	208-96-8	di-n-octyl phthalate	117-84-0
aniline	62-53-3	diphenylamine	122-39-4
anthracene	120-12-7	2,3,5,6-tetrachlorophenol	935-95-5
azobenzene	103-33-3	fluoranthene	206-44-0
benzo[a]anthracene	56-55-3	fluorene	86-73-7
benzo[b]fluoranthene	205-99-2	hexachlorobenzene	118-74-1
benzo[k]fluoranthene	207-08-9	hexachlorobutadiene	87-68-3
benzo[ghi]perylene	191-24-2	hexachlorocyclopentadiene	77-47-4
benzo[a]pyrene	50-32-8	hexachloroethane	67-72-1
benzyl alcohol	100-51-6	indeno[1,2,3-cd]pyrene	193-39-5
bis(2-chloroethoxy) methane	111-91-1	isophorone	78-59-1
bis(2-chloroethyl) ether	111-44-4	2-methyl-4,6-dinitrophenol (wet)	534-52-1
bis(2-chloro-1-methylethyl) ether	108-60-1	1-methylnaphthalene	90-12-0
4-bromophenyl phenyl ether	101-55-3	2-methylnaphthalene	91-57-6
butyl benzyl phthalate	85-68-7	2-methylphenol	95-48-7
carbazole	86-74-8	3-methylphenol	108-39-4
4-chloroaniline	106-47-8	4-methylphenol	106-44-5
4-chlorodiphenyl ether	7005-72-3	naphthalene	91-20-3
4-chloro-3-methylphenol	59-50-7	2-nitroaniline	88-74-4
2-chloronaphthalene	91-58-7	3-nitroaniline	99-09-2
2-chlorophenol	95-57-8	4-nitroaniline	100-01-6
chrysene	218-01-9	nitrobenzene	98-95-3
dibenz[a,h]anthracene	53-70-3	2-nitrophenol	88-75-5
dibenzofuran	132-64-9	4-nitrophenol	100-02-7
di-n-butyl phthalate	84-74-2	N-nitrosodimethylamine	62-75-9
1,2-dichlorobenzene	95-50-1	N-nitrosodi-n-propylamine	621-64-7
1,3-dichlorobenzene	541-73-1	pentachlorophenol	87-86-5
1,4-dichlorobenzene	106-46-7	phenanthrene	85-01-8
2,4-dichlorophenol	120-83-2	phenol	108-95-2
diethyl phthalate	84-66-2	pyrene	129-00-0
2,4-dimethylphenol	105-67-9	pyridine	110-86-1
dimethyl phthalate	131-11-3	2,3,4,6-tetrachlorophenol	58-90-2
1,2-dinitrobenzene	528-29-0	1,2,4-trichlorobenzene	120-82-1
1,3-dinitrobenzene	99-65-0	2,4,5-trichlorophenol	95-95-4
1,4-dinitrobenzene	100-25-4	2,4,6-trichlorophenol	88-06-2
2,4-dinitrophenol	51-28-5	dioctyl phthalate	117-81-7
2,4-dinitrotoluene	121-14-2	bis(2-ethylhexyl) adipate	103-23-1

EPA METHOD 8270

Quality control products

Semi Volatile (BNA) Matrix Spike Mix (11 components)		Product Number	Size
2000 µg/mL in Methylene Chloride		DRE-GA09000422DI	1 mL
		DRE-GS09000422DI	5 x 1 mL
5000 µg/mL in Methylene Chloride		DRE-GA09000423DI	1 mL
		DRE-GS09000423DI	5 x 1 mL
10,000 µg/mL in Methylene Chloride		DRE-GA09000424DI	1 mL
		DRE-GS09000424DI	5 x 1 mL
acenaphthene	83-32-9	N-nitrosodi-n-propylamine	621-64-7
4-chloro-3-methylphenol	59-50-7	pentachlorophenol	87-86-5
2-chlorophenol	95-57-8	phenol	108-95-2
1,4-dichlorobenzene	106-46-7	pyrene	129-00-0
2,4-dinitrotoluene	121-14-2	1,2,4-trichlorobenzene	120-82-1
4-nitrophenol	100-02-7		

EPA Method 8270 B/N Calibration Check Mix (7 components)		Product Number	Size
2000 µg/mL in Methylene Chloride		DRE-GA09000434DI	1 mL
acenaphthene	83-32-9	diphenylamine	122-39-4
benzo[a]pyrene	50-32-8	fluoranthene	206-44-0
1,4-dichlorobenzene	106-46-7	hexachlorobutadiene	87-68-3
di-n-octyl phthalate	117-84-0		

EPA Method 8270 Acid Calibration Check Mix (6 components)		Product Number	Size
2000 µg/mL in Methylene Chloride		DRE-GA09000435DI	1 mL
4-chloro-3-methylphenol	59-50-7	pentachlorophenol	87-86-5
2,4-dichlorophenol	120-83-2	phenol	108-95-2
2-nitrophenol	88-75-5	2,4,6-trichlorophenol	88-06-2

EPA Method 8270 Internal Standard Mix (7 components)		Product Number	Size
2000 µg/mL in Methylene Chloride		DRE-GA09000428DI	1 mL
		DRE-GS09000428DI	5 x 1 mL
4000 µg/mL in Methylene Chloride		DRE-GA09000429DI	1 mL
		DRE-GS09000429DI	5 x 1 mL
acenaphthene-d10	15067-26-2	naphthalene-d8	1146-65-2
chrysene-d12	1719-03-5	perylene-d12	1520-96-3
1,4-dichlorobenzene-d4	3855-82-1	phenanthrene-d10	1517-22-2
1,4-dioxane-d8	17647-74-4		

EPA METHOD 8270

Color changing surrogates for pH indication in aqueous semi-volatile extractions EPA Method 8270 (BNA)

If you have ever calculated the expense and the level of frustration caused by a missed pH adjustment, you can understand the importance of getting it right the first time. With Color Changing Surrogates, the guess

work of wondering “did I change the pH” is removed. Base/Neutral samples are yellow and acidified samples are pink. Color change occurs between pH 3.4 – 4.8.

Color Changing BNA Surrogates (7 components)		Product Number	Size
pH Indicating 100 µg/mL in Methanol		DRE-GA09000430ME	25 mL
2-fluorobiphenyl	321-60-8	pyrene-d10	1718-52-1
2-fluorophenol	367-12-4	p-terphenyl-d14	1718-51-0
nitrobenzene-d5	4165-60-0	2,4,6-tribromophenol	118-79-6
phenol-d5	4165-62-2		

BNA Surrogates (6 components)		Product Number	Size
pH Indicating 100/200 µg/mL in Methanol		DRE-GA09000446ME	
2-fluorobiphenyl	321-60-8		100 µg/mL
2-fluorophenol	367-12-4		200 µg/mL
nitrobenzene-d5	4165-60-0		100 µg/mL
phenol-d5	4165-62-2		200 µg/mL
p-terphenyl-d14	1718-51-0		100 µg/mL
2,4,6-tribromophenol	118-79-6		200 µg/mL
pH Indicating 100/150 µg/mL in Methanol		DRE-GA09000447ME	
2-fluorobiphenyl	321-60-8		100 µg/mL
2-fluorophenol	367-12-4		150 µg/mL
nitrobenzene-d5	4165-60-0		100 µg/mL
phenol-d5	4165-62-2		150 µg/mL
p-terphenyl-d14	1718-51-0		100 µg/mL
2,4,6-tribromophenol	118-79-6		150 µg/mL

BNA Surrogates (7 components)		Product Number	Size
1000 µg/mL in Methylene Chloride		DRE-GA09000431DI	1 mL
		DRE-GS09000431DI	5 x 1 mL
5000 µg/mL in Methylene Chloride		DRE-GA09000432DI	1 mL
		DRE-GS09000432DI	5 x 1 mL
		DRE-GA09000433DI	5 mL
2-fluorobiphenyl	321-60-8	pyrene-d10	1718-52-1
2-fluorophenol	367-12-4	p-terphenyl-d14	1718-51-0
nitrobenzene-d5	4165-60-0	2,4,6-tribromophenol	118-79-6
phenol-d5	4165-62-2		

EPA METHOD 8270

Quality control standards

EPA Method 8270 Semi-volatile System Performance Check Mix (4 components)		Product Number	Size
2000 µg/mL in Methylene Chloride		DRE-GA09000436DI	1mL
2,4-dinitrophenol	51-28-5	4-nitrophenol	100-02-7
hexachlorocyclopentadiene	77-47-4	N-nitrosodi-n-propylamine	621-64-7

Tuning

PFTBA MS Tuning Compound		Product Number	Size
(Perfluorotributylamine)	311-89-7	DRE-C15987500 neat	100 mg
1000 µg/mL in P/T Methanol		DRE-GA09010390ME	1 mL

GC Tuning Mix (4 components)		Product Number	Size
500 µg/mL in Methanol		DRE-GA09000425ME	1 mL
		DRE-GS09000425ME	5 x 1 mL
benzidine	92-87-5	decafluorotriphenylphosphine (DFTPP)	5074-71-5
p,p'-DDT	50-29-3	pentachlorophenol	87-86-5

DFTPP SV Tuning Compound		Product Number	Size
(Decafluorotriphenylphosphine)	5074-71-5	DRE-GA09010393AC	1 mL
1000 µg/mL in Acetone		DRE-GS09010393AC	5 x 1 mL
2500 µg/mL in Acetone		DRE-GA09010394AC	1 mL
		DRE-GS09010394AC	5 x 1 mL

For further information, please contact one of our local sales offices.



1 BRAZIL
T +55 12 3302 5880
E bz@lgcstandards.com

2 BULGARIA
T +359 (0)2 971 4955
E bg@lgcstandards.com

3 CHINA
T +86 400 9216156
E infochina@lgcgroup.com

4 FRANCE
T +33 (0)3 88 04 82 82
E fr@lgcstandards.com

5 GERMANY
T +49 (0)281 9887 0
E de@lgcstandards.com

EXPORT QUERIES
T +49 (0) 281 9887 250
E global.sales@lgcgroup.com

6 HUNGARY
T +40 364 116890
E ro@lgcstandards.com

7 IRELAND
T +44 (0)20 8943 8480
E uksales@lgcstandards.com

8 ITALY
T +39 02 22476412
E it@lgcstandards.com

9 MIDDLE EAST
T +49 (0) 281 9887 0
E fr@lgcstandards.com

10 NETHERLANDS
T +49 (0)281 9887 250
E nl@lgcstandards.com

11 NORDIC COUNTRIES
T +49 (0)281 9887 0
E de@lgcstandards.com

12 POLAND
T +48 22 751 31 40
E pl@lgcstandards.com

13 ROMANIA
T +40 364 116890
E ro@lgcstandards.com

14 RUSSIA
T +7 812 777 04 88
E ru@lgcgroup.com

15 SOUTH AFRICA
T +27 (0)11 466 4321
E sales.za@lgcgroup.com

16 SPAIN
T +34 (0)93 308 4181
E es@lgcstandards.com

17 TURKEY
T +90 530 880 33 90
E tur@lgcstandards.com

18 UK REFERENCE MATERIALS
T +44 (0)20 8943 8480
E uksales@lgcstandards.com

PROFICIENCY TESTING
T +44 (0) 161 762 2500

19 USA + CANADA
T +1 (603) 622 7660
E lgcusa@lgcgroup.com

For full listing of offices and distributors
please see lgcstandards.com





Order online at lgcstandards.com

No part of this publication may be reproduced or transmitted in any form or by any means, electronic or mechanical, including photocopying, recording or any retrieval system, without the written permission of the copyright holder. © LGC Limited, 2017. All rights reserved. updated 08/11/17