

Product PN **MV32ANPNY002BC01**
MV32ANPNY004UC01

Mod. 984

Rev. 02

Description **SYRINGELESS FILTER**

SEPARA (NY membrane) Minivial autosampler



PRODUCT DESCRIPTION	Syringeless filter that integrates in one single device auto-sampler vial, filtration membrane, plunger, and cap/septa
MANUFACTURER NAME	GVS S.p.A.
INTENDED USE / APPLICATION	Syringeless filter which provide a NY Hydrophilic - Low protein binding - Superior strength - Resistant to organic solvents. Compound Class: Bases, HPLC solvents, alcohols, aromatic hydrocarbons
CLASS OF THE PRODUCT	Laboratory Liquid Filtration
MATERIALS	<p>Filter media: NY hydrophilic membranes with 0.20 µm and 0.45 µm pore sizes</p> <p>Frame/Housing Polymer: Polypropylene</p> <p>Color: Transparent</p> <p>Regulatory Documentation Required:</p> <ul style="list-style-type: none"> <input type="checkbox"/> Biocompatibility according ISO 10993-1 <input type="checkbox"/> IMDS <input checked="" type="checkbox"/> Rohs, directive 2002/95/CE <input type="checkbox"/> BSE/TSE, directive 2003/32/CE <input type="checkbox"/> DEHP plasticizer Free and Latex free <input type="checkbox"/> Aging <input checked="" type="checkbox"/> Reach 1907/2006/CE (hazardous substances regulation) <input type="checkbox"/> Dir. 67/548/CE and Reg. 1272/2008/CE (medical sector dangerous substances) <input checked="" type="checkbox"/> Conflict minerals free

Product PN **MV32ANPNY002BC01**
MV32ANPNY004UC01

Mod. 984

Rev. 02

Description **SYRINGELESS FILTER**

PRODUCT CHARACTERISTIC	<p>Appearance/Visual:</p> <p>Aesthetic part: transparent body with colored cap Presence of graphics/Artwork: membrane type and pore size are pad printed on the colored cap Flash bears admitted: smaller than 0.3mm Injection gate vestige: smaller than 0.3mm Black spots admitted (size/area): must not be visible to the naked eyes at arm's length Infiltration into insert admitted: smaller than 25% of filtration area</p> <p>Physical/Mechanical:</p> <p>Dimensions: 12 x 32 mm Weight: 1.9 g Operating temperature Range: From 10 °C to 40 °C Storage temperature Range: From 10 °C to 40 °C</p> <p>Functional:</p> <p>Liquid Flow Rate with 0.20 µm : > 0.5 ml/min/cm² @ 80 cm H₂O Liquid Flow Rate with 0.45 µm : > 2.5 ml/min/cm² @ 80 cm H₂O Bubble Point Pressure with 0.20 µm : > 1.9 bar for 10 sec Bubble Point Pressure with 0.45 µm : > 0.9 bar for 10 sec Maximum volume 480 µL Dead Volume: < 30µL Operating Lifetime: 3 years Shelf Lifetime: 3 years</p> <p>Cleanliness:</p> <p>Clean-room Production required: Class ISO 8</p>
INSTRUCTIONS WARNINGS	<p>IFU: <input checked="" type="checkbox"/> English <input type="checkbox"/> French <input type="checkbox"/> Italian <input type="checkbox"/> Spanish <input type="checkbox"/> Portuguese <input type="checkbox"/> German <input type="checkbox"/> Chinese <input type="checkbox"/> Japanese <input type="checkbox"/> Korean <input type="checkbox"/> Russian <input type="checkbox"/> Other(Specify):</p>
STERILIZATION	<p>Not required</p>
PACKAGING AND LABELLING	<p>Packaging : External with the box with product logo [dim 170x80x120 mm] 2 Internal plastic bags with zip lock [dim 125x180 mm]</p> <p>Labelling: Standard Each box carton shall be labelled with the product part number and lot number. 100pcs per box . Different lots in one box are not allowed. IFU available (only when required by customer).</p>

PRODUCT SPECIFICATION

Product PN **MV32ANPNY002BC01**
MV32ANPNY004UC01

Mod. 984

Rev. 02

Description **SYRINGELESS FILTER**



CERTIFICATE OF COMPLIANCE

With each shipment, GVS S.p.A. Customer Service will send the CoC to the Customer, based on the lot numbers and date of manufacture.
 Conformity declaration is printed on every invoice and Certificate is according to UNI EN 10204 type 2.1
 The Quality management system is in compliance with ISO 9001

DRAWING

The attached drawing is part of this material specification and must not be duplicated or made accessible to a third party without prior written GVS SpA consent.

VISUAL REQUIREMENTS

Visual acceptance requirements apply when inspected under below conditions:

Magnification: unaided eye, approximately 45 cm (18") from eye
Illumination: 1000 ± 200 lx or equivalent
Light type: Fluorescent
Timings: 5 sec per unit

Acceptance Requirement		AQL	Sampling Plan
1	1. Incomplete / deformation of plastic support, shrinkages of material	1	ISO 2859 part.1 1 st Level
	2. Incomplete / missing / misplaced / unsealed membrane		
	3. Membrane broken, pierced, crushed, peeled away		
	4. Infiltration > 25% of material membrane surface		
	5. Membrane incorporated residual		
	6. No missing components or wrong color cap		
	7. Serious mold, rust dirty		
2	8. Flashes > 0.3 mm	4	
	9. Infiltration < 0.25% of the membrane surface		
	10. * Embedded particulate matter > 0.1 mm ²		
	11. Color not uniform		
	12. Marks or scratches		
	13. Loose particulate > 0.1 mm ² or fibers		

- Embedded particulate matter : according to Dirt Estimation Chart (Tappi Standard)

PRODUCT SPECIFICATION

Product PN **MV32ANPNY002BC01**
MV32ANPNY004UC01



Mod. 984

Rev. 02

Description **SYRINGELESS FILTER**

PERFORMANCE REQUIREMENTS	Acceptance Requirement		AQL	Sampling Plan	
	1	Bubble point test with 0.20µm pore size	≥ 1.9 bar (28 psig) for 10 sec	0.4	ISO 2859-1 Level S-3
	2	Bubble point test with 0.45µm pore size	≥ 0.9 bar (13 psig) for 10 sec	0.4	

REVISIONS AND APPROVALS:

DATE	REV.	REASON FOR CHANGE	ISSUED AND CONTROLLED BY: (name /function and signature)	APPROVED BY: (name /function and signature)
10.05.2017	00	First emission	Project Leader Simone Reggiani  Filippo Trascinelli	Quality Assurance Roberta Frulla 

Customer Approval:

We accept this material specification as a part of the agreed terms of delivery

Company name _____

Approved by: _____
 (Name, Function) (Signature)

Date _____
 (Company stamp)

Please send back this document signed for approval. If we will not receive this specification signed, we consider the first order placed as implicit approval.